

Flowline

SPECIALISTS IN FLOW MEASUREMENT

PERFORMANCE CHECK

CUSTOMER	Flowserve	CERTIFICATE NO.	05/06/2018 - 1655
MODEL NO.	MinisonicP+1595	CALIBRATION MEDIUM	WATER
SERIAL NO.	1655+8697&8698	DENSITY (KG/M3)	1000
ENGINEER	IC	VISCOSITY (CP)	1
DATE	05/06/2018		
TEMPERATURE	7	MASTER FLOWMETER	MAG5100W MAG6000
TEST RIG	AA	MASTER CERT	SN: 232102H405

TRACEABILITY

Master Meter is traceability to national standards.

MASTER FLOW M3/H	INDICATED FLOW M3/H	% ERROR OF READING	% OF F.S. FLOW RATE
252.6	252.8	0.08%	100.0%
223.9	225.5	0.71%	88.6%
198.7	201.1	1.21%	78.7%
176.1	176.7	0.34%	69.7%
151.0	152.5	0.99%	59.8%
126.4	127.4	0.79%	50.0%
102.0	102.1	0.10%	40.4%
74.9	75.9	1.34%	29.7%
52.1	51.6	-0.96%	20.6%
26.1	25.8	-1.15%	10.3%

SIGNED



Flowline Systems Ltd. Devonshire Business Centre, Works Rd, Letchworth Garden City, Herts, SG6 1GJ

Telephone 01462 488 224, Email Sales@FlowlineSystems.co.uk Website www.flowline.co.uk

Registered in England No. 5733073, Devonshire Business Centre, Works Rd, Letchworth Garden City, Herts, SG6 1GJ